

# Verification

of Test Report No.: IL130503010C1



**Product.....:** Digital Multimeter (CATII 600V & CATIII 300V)

**Model No.....:** 09DM1381

**Applicant.....:** King Tony Tools Co., Ltd.

No 11 , Alley 150, Lane 516, Sec. 2, Hsi Nan Rd., Wu-Jih Dist. 41467,  
Taichung City, Taiwan.

**Issue Date.....:** Mar.27, 2018

A sample of the equipment has been tested for CE marking  
according to the EC Low Voltage Directive 2014/35/EU.

Standard(s) used for showing compliance with essential requirements of the directive.

Standard:

EN 61010-1:2010 and EN 61010-2-030:2010 and EN 61010-031:2002 +A1:2008

Safety requirements for electrical equipment for measurement, control, and laboratory use-

Part 1: General requirements

Part 2-030: Particular requirements for testing and measuring circuits

Part 031: Safety requirements for hand-held probe assemblies for electrical measurement  
and test

The referred test report(s) show that the product fulfills the requirement in the EC Low Voltage Directive for CE marking. On the basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify with the EC Low Voltage Directive.

This verification shall be accompanied with test report.

Approved by:

(Managing Director)



Test site: Integrity EnE Lab Inc.

12F, No.27-1, Ln. 169, Kangning St., Xizhi Dist., New Taipei City 221, Taiwan. (R.O.C.)

Tel: +886-2-2695-6367 Fax: +886-2-8695-1713